

Welch Allyn® DECLARATION OF CONFORMITY

SAP DIR No.: 80017313

Version: A

We declare, under our sole responsibility, that the product listed below conforms to the provisions of European Council Directive 93/42/EEC of 14 June 1993 concerning medical devices.

Manufacturer's Name and Business Address: Welch Allyn, Inc.
4341 State Street Road
Skaneateles Falls, NY 13153

EC REP Regulatory Affairs Representative
Welch Allyn Limited
Navan Business Park
Dublin Road
Navan, County Meath
Republic of Ireland

Product Name: Connex Module Accessories

REF 405672 (NIBP)
405701 (SureTemp Thermometer)
405712 (Nellcor SpO2)
408186 (SureTemp Thermometer)

Annex: II

Classification: IIa

Classification Rules: 10

GMDN Code and Term: 36551 – Patient monitoring system module, blood pressure, noninvasive
36562 – Patient monitoring system module, temperature
36554 - Patient monitoring system module, pulse oximetry

UMDNS Code and Term 11753 - Noninvasive Blood Pressure (NIBP) Modules, Physiologic
11761 - Temperature Modules, Physiologic Monitor
11763 - Pulse Oximetry Modules, Physiologic Monitor

Notified Body: (CE 0297) DQS Medizinprodukte GmbH,
August-Schanz-Str.21, 60433 Frankfurt am Main
certificate 314505 MR2.

Standards Applied: EN 1060-1 Non-invasive sphygmomanometers - Part 1: General requirements
EN 1060-3 Non-invasive sphygmomanometers - Part 3: Supplementary requirements for electro-mechanical blood pressure measuring systems

SAP DIR No.: 80017313

Version: A

EN 60601-1	Medical electrical equipment - Part 1: General requirements for safety
EN 60601-1-2	Medical electrical equipment – Part 1-2: General requirements for safety – Collateral standard: Electromagnetic compatibility – Requirements and tests
EN 60601-1-4	Medical electrical equipment - Part 1-4: General requirements for safety - Collateral standard: Programmable electrical medical systems
EN 62304	Medical Device Software – Software Life-Cycle Processes
EN 62366	Medical devices – Application of usability engineering to medical devices
EN 9919	Medical Electrical Equipment – Particular Requirements for the Basic Safety and Essential Performance of Pulse Oximeter Equipment for Medical Use

Authorised Signatory:


Paul Oris, Regulatory Affairs Representative2013-07-25
Date

WelchAllyn® DECLARATION OF CONFORMITY

SAP DIR No.: 80017313

Version: A

Document Change History

Version	Description	Author	Date
A	Initial Release	S. Schmidt	2011-11-02